Contrast Warmer Temperature Log

Temperature Range of Warmer: 20 degrees C to 30 degrees C.

Alarm to Trigger when below 15 degrees C & above 37 degrees C.

Instructions:

- Technologist to check the temperature of the contrast warmer weekly.
- Do not store food or beverages in the warmer.
- Do not store other substances in the warmer unless stated by the manufacture of the substance.
- Notify supervisor and Biomed if there is an issue with the warmer.

Date *(DD-MMM-YYYY)*	**Current Temperature ©**	**Initial of MRT**	**Action/ Follow-up** *(when temperature is out of range OR warmer is not functioning properly)*